

Department of Chemistry
Dr. Hari Singh Gour University, Sagar, M.P.-470003

Ref. No.: Chem/SKS/UGCPROJ-Son/1/13

Dated: 10th April 2013

Quotations in sealed envelopes are required for the following equipment under UGC sponsored research project sanctioned to Prof. S.K. Srivastava, Department of Chemistry, Dr. Hari Singh Gour University, Sagar, M.P.-470003

1. Ultrasonic bath(Ultrasonicator)

Specification:

- Power: 500 Watts
- Output frequency : 20+3 KHZ
- Electronic control & temperature monitoring
- Sample processing volume: 5-900ml
- Automatic tuning set
- Input voltage: 220-240V A.C., 50 HZ
- Sound abating chamber

The quotations should reach to Prof. S.K.Srivastava, within fifteen days from the date of advertisement by Registered/Speed post.